



Hemoglobin A_{1c} measurement with Mono S method

Intellectual Property Notice: The Biopharma business of GE Healthcare was acquired by Danaher on 31 March 2020 and now operates under the Cytiva™ brand. Certain collateral materials (such as application notes, scientific posters, and white papers) were created prior to the Danaher acquisition and contain various GE owned trademarks and font designs. In order to maintain the familiarity of those materials for long-serving customers and to preserve the integrity of those scientific documents, those GE owned trademarks and font designs remain in place, it being specifically acknowledged by Danaher and the Cytiva business that GE owns such GE trademarks and font designs.

cytiva.com

GE and the GE Monogram are trademarks of General Electric Company.
Other trademarks listed as being owned by General Electric Company contained in materials that pre-date the Danaher acquisition and relate to products within Cytiva's portfolio are now trademarks of Global Life Sciences Solutions USA LLC or an affiliate doing business as Cytiva.
Cytiva and the Drop logo are trademarks of Global Life Sciences IP Holdco LLC or an affiliate.
All other third-party trademarks are the property of their respective owners.
© 2020 Cytiva
All goods and services are sold subject to the terms and conditions of sale of the supplying company operating within the Cytiva business. A copy of those terms and conditions is available on request. Contact your local Cytiva representative for the most current information.
For local office contact information, visit cytiva.com/contact

CY14246-08Jun20-AN

Hemoglobin A_{1c} measurement with Mono STM method

Abstract

Mono S 5/50 GL (TricornTM) columns are suitable for the measurement of HbA_{1c} and the calibration of healthcare instruments that analyze HbA_{1c}. The Mono S method is the Swedish designated method since it is effective in separating components that do not represent HbA_{1c}. With a greater resolution than both the Japanese and US designated methods, Mono S gives results that are closer to the HbA_{1c} standard set by the International Federation for Clinical Chemistry.

The performance of the Tricorn column, Mono S 5/50 GL, was compared with the older design Mono S HR 5/5 column for the analysis of HbA_{1c}. No differences were found between the results from the old and new columns. The Tricorn columns exhibited a high degree of precision in tests both within and between assays.

Diabetes and HbA_{1c}

Measurement of Hemoglobin-A_{1c} (HbA_{1c}) in diabetic patients is the most established procedure used for evaluating the long-term control of diabetes (1, 2). Hemoglobin-A_{1c} reflects the average blood glucose concentration for the past three months, which is the approximate life span of red blood cells. Hemoglobin, like many other proteins in blood, links up with sugars, such as glucose. In healthy individuals the HbA_{1c} peak represents 4%-5% of the total hemoglobin compared to 10%-15% in untreated diabetics. The goal of successful treatment is to keep the level of HbA_{1c} as close to the non-diabetic conditions as possible depending on the situation for the individual patient.

HbA_{1c} is a stable product where glucose is attached to the amino acid valine at the N-terminal of the β -chains of the molecule. Other possible glycation sites are the N-terminals of the α -chains as well as some lysine residues that can be glycosylated to some extent depending on the blood glucose

Column: Mono S 5/50 GL
 Sample: hemolyzed EDTA blood
 Sample volume: 10 μ l
 Buffer A: 20 mM sodium malonate, 0.2 g/l sodium azide, pH 5.7
 Buffer B: buffer A + 0.3 M LiCl
 Gradient: 20%-50% B for 3 min, 50%-75% for 1.1 min, 75%-100% for 0.3 min, 100% for 2.6 min, 20% for 1 min.
 Flowrate: 2 ml/min
 System: HPLC system
 Detection: 415 nm




Fig 1. Mono S ion exchange chromatography of HbA_{1c}. HbA3 represents the Hb-glutathione adduct and the HbF peak also contains some minor hemoglobins.

level. The glycation of the N-terminals changes the charge of the amino groups, which results in a clear separation with the Mono S method of the glycosylated (HbA_{1c}) and the non-N-terminal glycosylated form (HbA0). The hemoglobin molecule dissociates into α - β -dimers at pH 5.7. Thus the major content of the HbA_{1c} peak contains glycosylated- β -chain-non-glycosylated- α -chain dimer and a minor amount of non-glycosylated- β -glycosylated- α -chain dimer. A small amount, 0.3%, of carbamylated β -chains is also present during these separation conditions (Fig 1).



Mono S and calibration

Regular calibration of instruments used to measure HbA_{1c} is necessary due to the wide variety of instruments and techniques used in healthcare services. Since national standardization programs are based on arbitrarily chosen methods, there are problems in comparing results regionally and internationally. Therefore a methodologically sound, international reference measurement system has been developed by the International Federation for Clinical Chemistry (IFCC) for worldwide calibration (3).

Figure 2 shows how well three nationally designated comparison methods compare with the 'true' proportion of HbA_{1c} as defined by the IFCC. Many countries and instrument manufacturers calibrate their techniques with a method from the United States National Glycoprotein Standardization Program (NGSP [4]). The NGSP method overestimates the proportion of HbA_{1c} in blood samples since the chromatography used identifies a proportion of components that do not represent HbA_{1c}. The method recommended by the Japanese Diabetes Society (JDS), is more specific and lies closer to the IFCC standard (Fig 2). The Swedish designated comparison method is based on cation exchange chromatography and uses Mono S. This method has the highest resolution for the HbA_{1c} top and therefore lies closest to the standard set by the IFCC.




Fig 2. A comparison of HbA_{1c} measured with nationally designated methods and HbA_{1c} as defined by the IFCC method (5).

From HR to Tricorn

For almost two decades the Mono S method was conducted with Mono S HR 5/5 columns from GE Healthcare (6, 7). These columns were replaced with Mono S 5/50 GL in the Tricorn column design.

Tricorn columns incorporate features designed to meet the demands of today's laboratories, while retaining the standards set by HR columns. The principle modifications include Valco™ fittings for simple connection to ÄKTA™ design systems. A clear coating on the glass column protects users and the column without impairing visibility of the media bed. Tricorn columns have a smaller dead volume than HR columns and no fixed capillary tubing that allows optimization of capillary tube length and resolution. Finally, Tricorn columns have

fewer components than HR columns, making them robust and simple to operate and maintain. The analysis of HbA_{1c} on Tricorn columns has been compared with the old Mono S HR 5/5 at the Department of Clinical Chemistry, Malmö University Hospital, Sweden.

Materials and methods

Preparation of samples

Blood was collected in EDTA-containing tubes. 14 µl of the mixed whole blood was diluted and hemolysed with 700 µl of a buffer containing Citric acid, 0.02 mol/L, Na₂HPO₄(2H₂O), 0.055 mol/l and Triton™ X-100, 0.1% w/v, pH 5.4. After incubating the hemolysed cells at 37°C for 30 min to eliminate pre-HbA_{1c} the tubes were then centrifuged at 3000 × g for 10 min. 10 µl of the supernatant was injected into the column.

Columns and procedures

Analysis of HbA_{1c} was conducted on HPLC with following columns, Mono S 5/50 GL (Media lot nr. 286415) and Mono S HR 5/5 (Media lot nr. 288385). HbA_{1c} was separated from other hemoglobin forms by a LiCl-gradient. Buffer A contained 20 mmol/l of sodium malonate and 0.2 g/l of sodium azide, pH 5.7. Buffer B was buffer A plus 0.3 mol/l of lithium chloride. Before use, the buffers were degassed and filtered through a 0.45 µm pore filter. Flow rate was 2 ml/min. 10 µl was injected on the column.

Column equilibration

To equilibrate the columns for first time use before HbA_{1c} analysis a number of injections are necessary before the chromatography yields acceptable results. Forty samples were loaded and injected overnight to dispose of tailing caused by the HbA₀ fraction.

Results and Discussion

To determine whether Mono S 5/50 GL (Tricorn) columns were comparable with Mono S HR 5/5, 30 blood samples from diabetic patients and non-diabetics were assayed for HbA_{1c} (Fig 3). Assays with the Tricorn column had a near-perfect correlation to assays with the Mono S HR 5/5 column ($R^2 = 0.999$).




Fig 3. Comparison of HbA_{1c} separated with Tricorn and HR columns.

The within-assay precision was examined with a Tricorn column. This was done by repeatedly assaying HbA_{1c} in two series of 21 samples from a non-diabetic and diabetic patient. The results revealed high reproducibility as indicated by the low coefficient of variance (CV) for assays (Table 1).

Table 1. Precision within assay

Within Assay	n	Mean	SD	CV (%)
Non-diabetic	21	5.0	0.02	0.40
Diabetic	21	8.9	0.02	0.27

Between-assay precision with a Tricorn column was assessed by repeatedly assaying samples from three test groups with low, medium, and high HbA_{1c} content over a five-day period. The results all had CV values that were < 1.5% (Table 2), indicating a good analytical procedure on a well-maintained instrument. The higher CV values found in the between-assay test (Table 2) when compared to the within assay test (Table 1) were probably due to changing reagents between the assay runs.

Table 2. Precision between assays

Between Assay	n	Mean	SD	CV (%)
Low	15	5.0	0.08	1.50
Medium	25	7.0	0.08	1.14
High	15	8.9	0.09	1.05

Conclusion and discussion

The Mono S method is a reliable and well-tested method for the measurement and calibration of HbA_{1c} instruments. The method is currently used by network of accredited laboratories that are responsible for calibration of all Swedish hospital and point of care equipment independent of manufacturers (8).

A comparison of Tricorn Mono S 5/50 GL with Mono S HR 5/5 columns using the Mono-S method showed no noticeable effect on the quality of HbA_{1c} results as indicated by the high degree of correlation. The results confirm Tricorn columns have the same high chromatographic performance, stability and reproducibility as HR columns.

References

1. The Diabetes Control and Complications Trial Research Group. The effect of intensive treatment of diabetes on the development and progression of long term complications in insulin-dependent diabetes mellitus. *N.Engl.J.Med.* **329**, 977–986 (1993).
2. Stratton, I.M. *et al.* Association of glycaemia with macrovascular and microvascular complications of type 2 diabetes (UKPDS 35): prospective observational study. *B. M. J.* **321**, 405–411 (2000).
3. Jeppsson, J-O. *et al.* Approved IFCC Reference Method for Measurement of HbA_{1c} in Human Blood. *Clin. Chem. Lab. Med.* **40**, 78–89 (2002).
4. Little, RR. *et al.* The National Glycohemoglobin Standardization Program: A Five-Year Progress Report. *Clin. Chem.* **47**, 1985–92 (2001).
5. Hoelzel, W. *et al.* IFCC Reference System for Measurement of Hemoglobin A_{1c} in Human Blood and the National Standardization Schemes in the United States, Japan and Sweden: A Method-Comparison Study. *Clin.Chem.* **50**, 166–174 (2004).
6. Jeppsson, J-O. *et al.* Measurement of hemoglobin A_{1c} by a new liquid chromatographic assay: methodology, clinical utility and relation to glucose tolerance evaluated. *Clin. Chem.* **32**, 1867–72 (1986).
7. Eckerbom, S. *et al.* Improved method for analysis of glycated hemoglobin by ion exchange chromatography. *Ann. Clin. Biochem.* **31**, 355–360 (1994).

Acknowledgement

We would like to thank B. Larsson and Ass. Prof. J-O Jeppsson, at the Dept. of Clinical Chemistry, Malmö University Hospital, Malmö, Sweden for the information about comparisons of nationally designated methods and for comparing Tricorn and HR columns.

Ordering information

Article	Code No.
Mono S 5/50 GL	17-5168-01

For local office contact information, visit
www.gelifesciences.com/contact

www.gelifesciences.com/protein-purification

GE Healthcare Bio-Sciences AB
Björkgatan 30
751 84 Uppsala
Sweden



imagination at work

GE, imagination at work, and GE monogram are trademarks of General Electric Company.

ÄKTA, Mono S, Tricorn, and Drop Design are trademarks of GE Healthcare companies.

The Tricorn column and components are protected by US design patents USD500856, USD506261, USD500555, USD495060 and their equivalents in other countries.

All third party trademarks are the property of their respective owners.

All goods and services are sold subject to the terms and conditions of sale of the company within GE Healthcare which supplies them. A copy of these terms and conditions is available on request. Contact your local GE Healthcare representative for the most current information.

© 2004-2009 General Electric Company - All rights reserved.
First published June 2004

GE Healthcare UK Limited
Amersham Place, Little Chalfont, Buckinghamshire, HP7 9NA, UK

GE Healthcare Europe, GmbH
Munzinger Strasse 5, D-79111 Freiburg, Germany

GE Healthcare Bio-Sciences Corp.
800 Centennial Avenue, P.O. Box 1327, Piscataway, NJ 08855-1327, USA

GE Healthcare Bio-Sciences KK
Sanken Bldg., 3-25-1, Hyakunincho, Shinjuku-ku, Tokyo 169-0073, Japan